

# Order Form/Design Configuration for the Positioning Systems Frame Adacta Liga Light



Order 

Quotation

Invoicing Details	Shipping Details
Company name:	Ship to:
Delivery date required://	<del></del>
Delivery address:	CUSTOMER STAMP AND SIGNATURE
P. Code.: Town:	
Tel.: Fax:	
# Order: Date of the Order:/	

**Pro Medicare srl** grants **24 months** warranty on production defects from the date of the first put into service and **12 months** on covers and components subjected to wearing. Warranty is void if the product is used in any other way than stated in the Instruction Manual.

**Note:** Any request for variation in dimensions, materials and type compared to the standard does not have **CE** marking, therefore the professional user becomes the manufacturer and has the obligation to draw up the technical documentation of the device (this is valid for the sections of the sheet with the letter **W**; the prices of the **W** items are not present on this order form)

Images are purely illustrative and may not fully reflect reality.

## **Standard Configuration**

- Adacta Liga satin anodised Frame
- Fixed folding down back posts
- Reclining mechanism by predetermined plate w/pin 3°, 9°, 15°, 22°
- Continuous Tilt in space by drive lever:

paediatric from 2° up to 35° (from 6° up to 39°) 1

adult from 4° up to 30° (from 8° up to 34°) $^1$  tilt in space can be up-graded of 4° with increment of the starting inclination of the frame by operating on the positioning holes of the casters forks

- Fixed anti-tip system
- Wheels Kit with brake
- Adjustment of the seat depth
- Height adjustment: high (49 cm from the floor)/low (43 cm from the floor)\*

□ Rigid

- \*(with 300mm and 500mm wheels only)
- Knee angle
- Foot Plates
- Simple Calf Strap

Fra	me	Type									
114	1110	Type		□ Folding							
Ero	<b></b>	Colo		O Blue	O Red	Turquoise					
<u>FI a</u>	me	COIO	-	O Pink	0 Titan	ium					
Fra	me	Size									
			ts are from tui Ird engagemei	be to tube exterr nt).	nally, while de	pth measui	rements o	re from			
			Seat V	Vidth (cm)	/idth (cm) Depth and Adjustme						
<u>.</u>	0	US		30	0 36	◊ 38.5	0 41	0 44			
atr	0	XXS		35	0 36	◊ 38.5	0 41	0 44			
Paediatric	0	XS		35	0 40	◊ 42.5	0 45	0 48			
۵	0	XS0		38	0 40	◊ 42.5	0 45	0 48			
	0	XS1		38	0 44	◊ 46.5	0 49	0 52			
1			1		1						

044 **46.5** 049 052

048 **\$50.5** 053 056

048 **\$50.5** 053 056

044 **46.5** 049 052


048 050.5 053 056

41

41

44

47



- H: seat rails to floor height (see page 6)
- **Q:** length from seat rails to footplate (see page 6)
- C: push handles height to floor = 110 cm
- M: push handles height from seat rails (see page 2)

**Notes:** If a selection is not made, the product will be delivered in its standard configuration. (See **bold** items)

- **O** = **Standard** O = Optional without up-charge
- ☐ = Optional with up-charge

Mod.Ord.Ada.Lig.Lig. S.P. EN Rev.10 10/2023

S

M

L

XL

0 M1

FRAME OPTIONS									
Castors									
<b>◊</b> 150mm (6") po	oly grey O	black	175mm (7") poly O <i>grey</i>	O black					
Posterior Wheel	7 3 -7	2.0.0	273 (7 7 po.) - g.e)	- 2.00.					
	_	ro lover							
	y grey with parking bral y black with parking bra								
	umatic <i>grey</i> with parking								
	y black with drum brake	<u>-</u>							
	y grey with drum brake								
	umatic grey with drum	brake							
	y black with drum brake		ver						
	grey with drum brake	_							
	umatic grey with drum	_							
	y black with parking bra								
□ 400mm (16") poly	y black with drum brake	(only 45 cm height)							
□ 400mm (16") poly	y black with drum brake	and parking brake le	ver (only 45 cm height)						
□ 500mm (20") poly	grey without handrims	s with parking brake I	ever	. (001)					
□ 500mm (20") pne	umatic <i>grey</i> without ha	ndrims with parking b	□ handrims for 50 prake lever	0mm (20") wheels					
□ 500mm (20") pne	umatic <i>grey</i> with handr	ims with drum brake							
□ 500mm (20") pne	eumatic <i>grey</i> with handr	ims with drum brake	and parking brake lever						
			ce lever (not available on size US)						
	e is reduced to 20°, only 47	cm height)							
Knee Angle	^								
			at the same of the						
<b>Paediatric</b> Sizes									
from US-XS0		H		In the second second					
(detachable)			Anala 75°	Elevating (with teethed mechanism)					
	Angle 9		Angle 75°	from 80° up to 180° with a gap of 14°					
		S LH	ORH OLH	□ RH □ LH					
				99					
Adult Sizes			\\ وي	· • • • 8					
from XS1-XL (swing-away			H						
and	\ 3		7	Vertical Telescopic and Elevating					
detachable)	\ d	 	Elevating (with a spring ping) 6 positions: from 110° up to 160° with 12°	6 positions: from 110° up to 160° with 12° gap					
	Angle 75°	Angle 90° O RH O LH		RH LH  Note: with knee joints padded cover					
Footrest		3 1111 3 211		Total Mar Mice Jemes padded cover					
♦ footplates	O footboard	□ slotted (foot	tplates only for intra-extra rotation of 15	;°)					
padding and cover				. ,					
Back Posts	in leather flex	ded - one	piece						
-	. Is a seall a c								
• Kit for fixed push		by nin (may adjustm	ent 10 cm with 2.5 cm gap)						
· ·	sh handles equal to 63 cm	by pill (max aujustill	ent 10 cm with 2.5 cm gap)						
☐ Kit for telescopic	push handles with leve	r block (max adjustm	ent 10 cm)						
☐ Kit for angle adju	istable single handlebar	for rigid frame							
		_	e (only for paediatric frame sizes)						
	and angle adjustable si	_							
-	ns for easier backrest in	clination (not valid	for crash test)						
<u>Tilt in space</u>									
♦ with lever	O with pedal								
Anti-tip system									
<b>◊</b> fixed									
□ detachable									

Mod.Ord.Ada.Lig.Lig. S.P. EN Rev.10 10/2023

#### FRAME ACCESSORIES <u>Seat Base</u> Note: The wooden seat bases present posterior notches of 6 cm for an easy movement on the seat rails □ wooden seat base covered with leather flex black fabric (with male and female hook and loop fastening straps) □ wooden seat base with mounting kit ☐ fixed grey wooden seat base with slots Abductor wedge only with rigid seat base Size (cm) $\Box$ S $\sqcap$ M $\sqcap L$ $\sqcap \mathbf{w}$ □ removable with lever □ removable mounting block 7.5 9 10 9 10 11.5 Note: if not indicate the abductor wedge will be removable with lever 7.5 9 L 10 □ standard Please indicate the position of holes under the seat (if not indicated the abductor wedge will be fixed in the $\hfill\Box$ on the RH not centred standard position □ on the LH **Vehicle Transportation Kit** (according to the ISO 7176-19 standard) Do not use the wheelchair as a seat in a vehicle if the weight of the user is lower than 22 kg □ only frame height 43 cm and with 300mm (12") and 500mm (20") posterior wheels (it is strongly recommended the use of a headrest ) Wooden Supporting Base for Breathing Equipment covered in black leather flex (available with tilt in space with lever only) **Note:** Bulky equipments may lead to a reduction of the tilt in space /reclining values □ with following maximum dimensions of the Breathing Equipment: Width: 25cm; Length: 30cm; Height: 10cm ☐ with Breathing Equipment maximum dimensions exceeding 25x30x10cm In this situation, please indicate: Width \_\_\_\_\_; Length \_\_\_\_\_; NOTE \_\_\_\_\_; NOTE \_\_\_\_\_ Note: Pro Medicare Srl reserves the right to check wheelchair compatibility **Pedal for barriers** Side $\square$ RH □ LH **Leg Support Separate Leg Cradle** (adjustable in height and depth) $\square$ RH $\Box$ LH □ custom W cm US XXS XS XS0 XS1 М1 L ΧL $\mathbf{W}_{\mathrm{RH}}$ $\mathbf{W}_{\mathsf{LH}}$ 10 12,5 12,5 15 15 15 18 18 18 9.5 12.5 12.5 12.5 12.5 14 14 15.5 17.5 17.5 ☐ **Calf Panel** (adjustable in height and depth) □ custom W XS1 ΧL US XXS XS XS0 S М M1 L W 10 12 12 15 15 18 18 18 18 18 28 23 26,5 26,5 25 25 28 31 34 34 □ Padded Calf Strap □ custom **W** US XXS XS XS0 XS1 S М M1 L ΧL w cm Н 11 11 11 11 11 13 13 13 13 13 34 37 37 40 43 L 26 31 31 34 43 Padded Divided Leg Cradles (adjustable in height and depth) $\square$ RH □ LH □ custom **W** $\mathbf{W}_{\text{LH}}$ XXS XS0 XS1 М1 ΧL $\mathbf{W}_{\mathrm{RH}}$ cm US XS S М L 17.5 17.5 12.5 15 15 17.5 20 20 20 ...... . . . . . . . . 15.5 15.5 1 12.5 14 14 17 17 19 22.5 22.5 13.5 14 15 15 15 15.5 15.5 16 16 ...... ...... $\hfill \Box$ Padded foot box (adjustable in height and depth) □ custom W XXS XS XS0 XS1 S M M1 L W н 12.5 15 15 17.5 17.5 17.5 20 20 22.5 22.5 23.5 29 29 32.5 32.5 35.5 35.5 38.5 41.5 41.5

15

15

**Cover for legrest tube** 

15

15.5

□ RH □ LH

......

3/7

13.5

**Cover for Knee Angle joint** □ RH □ LH

Mod.Ord.Ada.Lig.Lig. S.P. EN Rev.10 10/2023

# **Armrest** ☐ Height adjustable flip back tubular armrests/inclination max 5° o small (L = 25 cm) o large (L = 30 cm) ☐ Kit of ABS laterals multi-adjustable, detachable and w/multi-adjustable arm support in width and depth with 2.5 cm gap make a selection: make a selection: O small arm support (from US to XS0) O short vertical bar (from 22 cm to 31 cm\*) (dimensions: 5 cm x 25 cm) O long vertical bar (from 29 cm to 38 cm\*) O large arm support (from XS1 to XL) \* distance from armrest to the seat rail (dimensions: 6 cm x 30 cm) □ Kit of hardware multi-adjustable, detachable for ABS laterals make a selection: make a selection: O hip guide pad ① see below O short vertical bar (from 19 cm to 28 cm\*) O long vertical bar (from 26 cm a 35 cm\*) O arm support pad ②see below \*distance from hardware to the seat rail ABS Lateral: □ RH □ LH Padding for ABS Laterals □ RH □ LH Spacer for ABS laterals □ 1 cm □ 1.5 cm □ 2 cm □ RH □ LH Hin Guides

nip di	<u>ilues</u>																		
□ Line	□ Linear Hip Guide ① (pad only)										□ Contoured Hip Guide ① (pad only)								
☐ Linear Hip Guide pad w/mounting bracket									☐ Contoured Hip Guide pad w/mounting bracket										
. ~	cm XS S M L W <sub>RH</sub> W <sub>LH</sub>							W <sub>LH</sub>		1	cm	XS	S	М	L	W <sub>RH</sub>	W <sub>LH</sub>		
		Н		8	10	10	12				5	Н	8	10	10	12			
<b>*</b>	= 2,5 cm	L		22	25	30	35			s= 2,5 cm		L	22	25	30	35			
Side	Pad size	•	Select the bracket*  Select the bracket shape*						Side   Pad size   Select the bracket*						ect the et shape*				
			⊒ Fi	xed				□A	□В	□ Fixed				□A	□В				
□ RH □ Detachable □ C □ D						□D	□ RH □ □ Detachable					□C	□D						
□ Fixed □ A □ B							□В	LH		□ Fix	ked				□A	□В			
☐ LH ☐ Detachable ☐ C ☐ D								□D			□ De	etachal	ole			□C	□D		
Bracket Padding: □ RH □ LH								Bracket	Padding:	□ R	H 🗆 I	_H			·				

Note: If the bracket feature is not selected the fixed bracket will be delivered; for the choice of bracket shape see below

# \*Brackets (They allow adjustments of the seat and/or backrest width)

#### Shape

- ${\bf A}\;\; L/90^{\circ}$  Bracket allows a reduction of the width up to 2.5 cm per side
- $\,{\bf B}\,$  2.5 cm Offset Bracket allows a reduction of the width up to 5 cm per side
- C 5 cm Offset Brackets allows a reduction of the width up to 7.5 cm per side
- **D** Reverse Bracket doesn't allow any variation of the width









#### **Type**







Detachable Bracket

Mod.Ord.Ada.Lig.Lig. S.P. EN Rev.10 10/2023

Arm Support	<u>s</u>															
□ Flat Arm Support ② (pad only) □ Inside/outside flat Contoured Arm Support											oort ②	) (pad oi	nly)			
☐ Flat arm su	pport w/	□ Inside/ou	ıtside '	flat Co	ontoure	d arm	supp	ort w	bracke	ts						
s=2,5	cm	XS	S	М	L	<b>W</b> <sub>RH</sub>	W <sub>LH</sub>			cm	xs	S	М	L	<b>W</b> <sub>RH</sub>	W <sub>LH</sub>
	> н	8	10	10	12				3	Н	8	10	10	12		
~~	L	22	25	30	35				/H	L	22	25	30	35		
Side	Size							Side	S	Size						
□ RH								□ RH								
□ LH								□ LH								
☐ Flat and Sh	aped Arm	n Suppo	ort ②	(pad d	only)			☐ Inside/outside Contoured and Shaped Arm Supports ② (pad only)								
☐ Flat and Sh	aped Arm	n Suppo	rt w/	brack	ets			☐ Inside/ou w/brackets	ıtside	Conto	oured ar	nd Sha	ped A	Arm Su	pports	
M	cm	xs	S	М	L	W <sub>RH</sub>	W <sub>LH</sub>			cm	xs	S	М	L	<b>W</b> <sub>RH</sub>	W <sub>LH</sub>
s=2,5 cm	Н	8	10	10	12					Н	8	10	10	12		
	L	30	35	40	48					L	30	35	40	48		
M 12 15 17 17 M 12 15 17 17																
Side	Тур	e of Sha	аре			Size		Side		Тур	oe of Sh	ape			Size	
□ RH	O palmar							□ RH	O palmar O elbow							
	O elbow O palmar									palmar						
O elbow																
Wooden Tray																
☐ Ergonomic			with	Hardy	ware	□ w/la	ateral fli	р			Ţ	ype of	hardw	are		
Ligonomic			WICH	· iui u	wa.c	□ deta	achable								7	
						□ w/la	ateral fli	р	,						Å	l
☐ Rectangula	r		with	Hard	ware	□ deta	chable			late	ral flip			det	achable	
Padding	□ black	leather	flex		blue lea	ther fle	x				<u> </u>			ucu	acriable	
□ Outer Edge																
	ergon	nomic tra	y dime	ension	าร					rectar	ngular tra	ay dim	ension	s		
		Size cm		US	□ P	□ <b>M</b>	□ <b>G</b>			•	Size cm		US	□ P	□М	□ <b>G</b>
	۵.	Р	4	17	53	58	66			۵	Р	4	15	50	55	63
· · · · ·	\ .°°	L	5	58	69	75	86	000	\		L		66	66	71	82
В		В	3	35	42	49	55	В			В	3	35	42	48	55
- L	-	Α	1	L7	19	21	24		-		Α	1	L7	19	21	24
Pelvic Belt (w/	'2 adjustn	nents V	ersa-F	ixatis	s) (for	other m	nodels p	lease see Fixati	s Range	e orde	r form)					
				m			□ XS		<b>□</b> S		[	<b>□</b> M			□ <b>L</b>	
			H (pad	heigh	it)		4.5		5.5			7			7	
Plastic buckl	Plastic buckle closure L (pad length) 12 17 22 27															
□ <u>Basket</u>																
□ Rain cover	□ Rain cover															
Foot Muff																
Foot Muff		3 Blue	□ P	ink	□ Gr	ey 🗆	Borde	aux				-				

Mod.Ord.Ada.Lig.Lig. S.P. EN Rev.10 10/2023 5/7

			FRAME	DIMENSI	ONS AND	FEATUR	ES			
Load (kg)/ Frame Weight* (kg)	US	xxs	xs	XS0	XS1	S	М	M1	L	XL
Rigid Frame	50/14.1	50/14.4	50/14.7	50/15.4	75/19.2	75/19.5	100/19.8	100/19.9	100/20.1	100/20.3
Folding Frame	50/14.9	50/15.3	50/15.6	50/15.9	75/19.5	75/19.9	75/20.2	75/20.3	75/20.5	75/20.8
* Weight of the frame, including knee angle and footplates as per standard configuration										

Overall d	Overall dimensions (cm) of the frame with backrest tubes (fixed) folded down, footrest and posterior wheels detached										
Size US XXS XS XSO XS1 S M M1 L XL											XL
Length		66	66	68	68	71	73	83	83	73	83
Width	С	30	30	30	30	30	32	32	32	32	32
wiath	R	42	48	48	51	51	54	54	57	54	60
Height		65	65	66	66	67	67	68	68	67	68
Mainht (lon)	R	11.4	11.7	12	12.7	12.9	13.1	13.5	13.6	13.8	14
Weight (kg)	С	12.2	12.6	12.9	13.2	13.3	13.6	13.9	14	14.2	14.5

**Note:** The length dimensions, referred to the average values, are subjected to the variations due to the displacement of the wheel plates. The values reported are intended for the standard configuration.

	Width (cm)											
		US	XXS	XS	XS0	XS1	S	М	M1	L	XL	
300mm wheel	standard	48	53	53	56	56	59	59	62	65	65	
300mm wheel	w/drum brake	49	54	54	57	57	60	60	63	66	66	
400/500mm	standard	50	55	55	58	58	61	61	64	67	67	
wheel	w/drum brake	53	58	58	61	61	64	64	67	70	70	
600mm wheel	/	/	55	55	58	58	61	61	64	67	67	
Pivot W	/idth (cm)	69	69	72	75	75	75	78	78	75	78	
Q*(min-max) F footpl	13-23	13-23	13-26	13-26	28-36	28-36	32-40	32-40	28-36	32-40		
	43-49 with 300/500mm wheels and 150/175mm castors											

H\*\* Seat rail to floor (cm)

43-49 with 300/500mm wheels and 150/175mm castors 45 with 400mm wheels and 150/175mm castors 47 with 600mm wheels and 150mm castors

Max Slope

**Note:** \*The value Q, referred to 90° footrest hanger, can be increased of about 5-7 cm by using the 75° hanger

<sup>\*\*</sup>The value H has been detected outside the seat tube at the plate of the tilt mechanism; the frame will be delivered with H=43 cm

	Lenght (cm)											
			US	XXS	XS	XS0	XS1	S	М	M1	L	XL
	Without Knee angle	L	70	70	73	73	75	75	78	78	75	78
300mm wheel	75° Knee Angle	L <sub>min</sub> /L <sub>max</sub>	92/99	92/99	93/100	93/100	97/104	97/104	100/107	100/107	97/104	100/107
	90° Knee Angle	L <sub>min</sub> /L <sub>max</sub>	88/96	88/96	91/99	91/99	93/100	93/100	96/103	96/103	93/100	96/103
	Without Knee angle	L	74	74	77	77	79	79	82	82	79	82
400mm wheel	75° Knee Angle	L <sub>min</sub> /L <sub>max</sub>	95/101	95/101	96/103	96/103	100/107	100/107	103/110	103/110	100/107	103/110
	90° Knee Angle	L <sub>min</sub> /L <sub>max</sub>	93/99	93/99	94/100	94/100	96/102	96/102	99/104	99/104	96/102	99/104
	Without Knee angle	L	78	78	81	81	84	84	87	87	84	87
500mm wheel	75° Knee Angle	L <sub>min</sub> /L <sub>max</sub>	98/103	98/103	99/107	99/107	103/113	103/113	106/116	106/116	103/113	106/116
	90° Knee Angle	L <sub>min</sub> /L <sub>max</sub>	95/101	95/101	97/102	97/102	100/106	100/106	103/109	103/109	100/106	103/109
	Without Knee angle	L	/	83	84	84	85	85	88	88	85	88
600mm wheel	75° Knee Angle	$L_{min}/L_{max}$	/	101/109	103/112	103/112	108/116	108/116	111/120	111/120	108/116	111/120
Wilcei	90° Knee Angle	L <sub>min</sub> /L <sub>max</sub>	/	98/104	101/107	101/107	105/112	105/112	108/116	108/116	105/112	108/116

L = Length of the frame without knee angle joint mechanism;

Mod.Ord.Ada.Lig.Lig. S.P. EN Rev.10 10/2023 6/7

 $L_{min}$  = Length of the frame with knee angle joint mechanism and footplate positioned backward;

 $L_{max}$  = Length of the frame with knee angle joint mechanism and footplate positioned forward

**Note:** The length dimensions, referred to the average values, are subjected to variations due to the displacement of the wheel plates. The values reported are intended for the standard configuration.

<u>Note</u>	
	_



### $\label{eq:promedicare s.r.l.} \textbf{PRO MEDICARE S.r.l.}$

Via Montagna Z.I. Lotto 41-72023 Mesagne (Br)-ITALY Tel: +39-0831-777840 Fax:+39-0831-730739 Website: <a href="www.promedicare.eu">www.promedicare.eu</a> e-mail: sales@promedicare.it